

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

MAY 12 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Datamaster cdm SN <u>130579</u> <del>X124669</del> ✓	
Name <u>Patrick E. Martin</u>	ID# <u>4072</u> Date <u>5-12-09</u> ✓
Agency <u>APD</u>	Phone # _____
Instrument Location <u>Jail</u>	
Alco S/N <u>X124669</u> ✓ Target Value <u>0.081</u> ✓ High Pressure <u>1175</u>	
Alco Test Value Average <u>0.075</u> ✓	
1 <sup>st</sup> Alco	<u>0.076</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>Patrick E. Martin</u> ✓	
COB 5/18/09	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

MAY 12, 2009 ✓

COB  
5/18/09

OPERATOR'S NAME:

MARTIN, PATRICK E

OPERATOR'S NUMBER: 4072

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

V

O.L. #: 1234567

DEPT/AGENCY: ANCL

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: X124669 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.11 in ✓

ALCO TARGET .081 13:48

BLANK TEST .000 13:49

INTERNAL STANDARD VERIFIED 13:49

ALCO TV 30.11 in .075 ✓ 13:50

BLANK TEST .000 13:50

SUBJECT SAMPLE .000 13:51

BLANK TEST .000 13:52

ALCO TV 30.11 in .076 ✓ 13:53

BLANK TEST .000 13:53

COB  
5/18/09

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

MAY 12, 2009 ✓

TIME 13:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.11 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefg  
hijklmnopqrstuvwxyz{|}~Δ