

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 15 2009

Supervisor/Operator Performing the Verification Procedure:	
Name <u>Derek Sitz</u>	DataMaster cdm S/N <u>130579</u> ✓
ID# <u>4096</u>	Date <u>5.14.09</u> ✓
Agency <u>Anchorage Police Department</u>	Phone # _____
Instrument Location <u>Anchorage Jail</u> ✓	
A	
Alco S/N <u>X 124669</u> ✓	Target Value <u>.081</u> ✓
	High Pressure <u>1150</u>
B	
Alco Test Values	
<u>.074</u> ✓	<u>.075</u> ✓
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>Derek Sitz</u> ✓	<u>COS</u> <u>5/18/09</u>
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

### NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
 DATAMASTER cdm 130579

MAY 14, 2009 ✓ *CSB 5/18/09*

OPERATOR'S NAME: *DS*  
 SITZ, DEREK W  
 OPERATOR'S NUMBER: 4096  
 SUBJECT'S LAST NAME: *1465*  
 VOC *5-14-09*  
 SUBJECT'S FIRST NAME/MI :  
 VOC  
 O.L. #: 1234567  
 DEPT/AGENCY: ANCI  
 CASE/REPORT: 00-1234  
 TEST TYPE: V  
 ALCO TARGET VALUE: .081 ✓  
 ALCO S/N: X124669 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.01 in		
ALCO TARGET	.081	00:38
BLANK TEST	.000	00:39
INTERNAL STANDARD	VERIFIED	00:39
ALCO TV 30.01 in	.074 ✓	00:40
BLANK TEST	.000	00:40
SUBJECT SAMPLE <i>CSB 5/18/09</i>	.000	00:41
BLANK TEST	.000	00:42
ALCO TV 30.03 in	.075 ✓	00:42
BLANK TEST	.000	00:43

### DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
 DATAMASTER cdm 130579

MAY 14, 2009 ✓  
 TIME 00:37

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY	05
PROGRAM:	OKAY	1465
SOFTWARE DATE:	12/11/08	5-14-09
HEATERS		
SAMPLE CHAMBER:	50c	
BREATH TUBE:	42c	
BAROMETER:	30.01 in	
FLOW DETECTOR:	OKAY	
PUMP		
HIGH SPEED:	OKAY	
DETECTOR:	OKAY	
FILTERS:	OKAY	
QUARTZ STANDARD:	OKAY	
CALIBRATION:	OKAY	

PRINTER TEST

```
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~`
```

