

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 13 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130579</u>
Name	<u>Derek Site</u>	ID#	<u>4096</u> Date <u>7-9-09</u>
Agency	<u>Anchorage Police Department</u>	Phone #	<u>786-8900</u>
Instrument Location	<u>Anchorage Jail</u>		
Alco S/N	<u>X124669</u>	Target Value	<u>.081</u> High Pressure <u>400 psi</u>
Alco Test Values	<u>.081</u> 1 st Alco	<u>.081</u> 2 nd Alco	
Signature	<u>Derek Site</u>		<u>003</u> <u>7/14/09</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of July, 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



LASKA DEPARTMENT OF PUBLIC SAFETY
ATAMASTER cdm 130579

JULY 09, 2009

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

.L. #: 1234567

EPT/AGENCY: ANCL

ASE/REPORT: 00-1234

EST TYPE: V

LCO TARGET VALUE: .081

LCO S/N: X124669

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.19 in		
LCO TARGET	.081	00:36
BLANK TEST	.000	00:36
INTERNAL STANDARD	VERIFIED	00:36
LCO TV 30.19 in	.081	00:37
BLANK TEST	.000	00:38
SUBJECT SAMPLE	.000	00:38
BLANK TEST	.000	00:39
LCO TV 30.19 in	.081	00:39
BLANK TEST	.000	00:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JULY 09, 2009

TIME 00:34

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.19 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklm
pqrstuvwxyz{|}~Δ