

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 14 2009

Supervisor/Operator Performing the Verification Procedure: Datamaster cdm S/N 130579 ✓

Name Patrick E. Martin ID# 4072 Date 8-12-09 ✓

A Agency APD Phone # _____

Instrument Location Jail - Rm # B104 ✓

B Alco S/N 77675 ✓ Target Value .082 ✓ High Pressure 1250

Alco Test Value Average .080 ✓ .082 ✓
1st Alco 2nd Alco

Signature [Signature] ✓

[Signature]
8/12/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130579 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

AUGUST 12, 2009

OPERATOR'S NAME:

MARTIN, PATRICK E

OPERATOR'S NUMBER: 4072

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

V

O.L. #: 1234567

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .082 ✓

ALCO S/N: 77675 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.17 in ✓

ALCO TARGET .082

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 30.17 in .080 ✓

BLANK TEST .000

SUBJECT SAMPLE ^{CAS} 8/28/09 .000

BLANK TEST .000 ✓

ALCO TV 30.17 in .082

BLANK TEST .000

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

AUGUST 12, 2009 ✓

TIME 15:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.16 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~△