

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130579 ✓</u>	
Name	<u>Troy Clark</u>	ID#	<u>4934 ✓</u> Date <u>09/07/2010 ✓</u>
A Agency	<u>Anchorage Police Department</u>	Phone #	<u>786-8900</u>
Instrument Location <u>Anchorage Jail</u>			
Alco S/N <u>2N ✓</u>		Target Value <u>.080 ✓</u>	High Pressure <u>1000</u>
Alco Test Values		<u>.080 ✓</u>	<u>.081 ✓</u>
		1 st Alco	2 nd Alco
Signature <u>[Handwritten Signature]</u>			

(OVER)

BMB
9/17/10


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 9th day of Nov, 2010.



[Handwritten Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

SEPTEMBER 07, 2010

OPERATOR'S NAME:

CLARK/TROY

OPERATOR'S NUMBER: 4934

SUBJECT'S LAST NAME:

CLARK

SUBJECT'S FIRST NAME/MI :

TROY

O.L. #: 6633880

DEPT/AGENCY: ANCL

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .080

ALCO S/N: 2N

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.06 in

ALCO TARGET	.080	02:06
BLANK TEST	.000	02:07
INTERNAL STANDARD	VERIFIED	02:07
ALCO TV 30.06 in	.080	02:08
BLANK TEST	.000	02:08
SUBJECT SAMPLE	.000	02:09
BLANK TEST	.000	02:10
ALCO TV 30.06 in	.081	02:10
BLANK TEST	.000	02:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

SEPTEMBER 07, 2010

TIME 01:58

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	30.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~Δ