

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

SEP 10 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130579</u>	
Name <u>S. Dunn</u>	ID# <u>4106</u> Date <u>9-9-09</u>
Agency <u>APD</u>	Phone # <u>786-8900</u>
Instrument Location <u>JAIL</u>	
Alco S/N <u>X301563</u> Target Value <u>.079</u> High Pressure <u>500 PSI</u>	
Alco Test Value Average <u>.080</u> <u>.079</u>	
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	
(OVER)	

[Signature]  
9/11/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 29th day of Sept, 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

SEPTEMBER 09, 2009

*BMB*  
*9/11/09*

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: V

DEPT/AGENCY: ANCI

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301563 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.67 in		
ALCO TARGET	.078	23:27
BLANK TEST	.000	23:28
INTERNAL STANDARD	VERIFIED	23:28
ALCO TV 29.67 in	.080 ✓	23:29
BLANK TEST	.000	23:29
SUBJECT SAMPLE	.000	23:30
BLANK TEST	.000	23:31
ALCO TV 29.68 in	.079 ✓	23:31
BLANK TEST	.000	23:32

*BMB*  
*9/11/09*

*S Dunn*  
*#1361*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

SEPTEMBER 09, 2009 ✓

TIME 23:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	29.67 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~Δ

*(S) #1361*