

FEB 08 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 01-28-08

**A** Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N X301532

Target Value .081

High Pressure 500


Alco Test Values:

**B**

.078  
1<sup>st</sup> Alco

.078  
2<sup>nd</sup> Alco

Signature \_\_\_\_\_

  
Supt. James R. R. Rea

(OVER)

*COB*  
*2/21/08*

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)


(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Rea, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed





Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

*02/22/08*

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130584

JANUARY 28, 2008

OPERATOR'S NAME:

REA, JAMES RR

OPERATOR'S NUMBER: 3161

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VERIFICATION OF CAL

O.L. #: 01282008

DEPT/AGENCY: OTZ1

CASE/REPORT: 01282008

TEST TYPE: V

ALCO TARGET VALUE: 081

ALCO S/N: X301532

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.65 in  
ALCO TARGET .000 11:46  
BLANK TEST .000 11:46  
INTERNAL STANDARD VERIFIED 11:46  
ALCO TV 29.64 in .078 11:47  
BLANK TEST .000 11:48  
SUBJECT SAMPLE .000 11:48  
BLANK TEST .000 11:49  
ALCO TV 29.64 in .078 11:49  
BLANK TEST .000 11:50

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130584

JANUARY 28, 2008

TIME 11:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.64 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""'()\*+,-./0123456789:;<=>@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~