

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161

Date 05-07-08

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 78890 ✓
~~X301532~~ CB

Target Value .082 ✓

High Pressure 1200

Alco Test Values:

.085 ✓
.083 ✓
5/29/08

1st Alco

.085 ✓
2nd Alco

B

Signature _____

Supt. James R. R. Rea
(OVER)

CB
5/31/08

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Rea, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

5/29/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

MAY 07, 2000

OPERATOR'S NAME:

REA, JAMES RR

OPERATOR'S NUMBER: 3161

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

VERIFICATION OF CAL

O.L. #: 05072000

DEPT/AGENCY: OTZ1

CASE/REPORT: 05072000

TEST TYPE: V

ALCO TARGET VALUE: 002

ALCO S/N: 70090

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.38 in
ALCO TARGET .003 09:21
BLANK TEST .000 09:22
INTERNAL STANDARD VERIFIED 09:22
ALCO TV 30.38 in .005 09:23
BLANK TEST .000 09:23
SUBJECT SAMPLE .000 09:24
BLANK TEST .000 09:24
ALCO TV 30.38 in .005 09:25
BLANK TEST .000 09:26

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

MAY 07, 2000

TIME 09:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS :
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 30.38 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJKL
MNPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrsstuvwxyz{|}~