

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 29 2010

DataMaster cdm S/N: 130584 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161 ✓

Date 07-25-10 ✓

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 78890 ✓

Target Value .082 ✓

High Pressure 0700

Alco Test Values:

.082 ✓

.082 ✓

B

1st Alco

2nd Alco

Signature _____

Supt. James R. R. Rea
(OVER)

CRS
7/30/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of AUGUST, 2010.

Daniel L. Nicholas
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JULY 25, 2010

OPERATOR'S NAME: *CBS 7/30/10*
REA/JAMES/RR
OPERATOR'S NUMBER: 3161
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 07252010
DEPT/AGENCY: OTZ1
CASE/REPORT: 07252010
TEST TYPE: V
ALCO TARGET VALUE: 002
ALCO S/N: 78890

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.07 in
ALCO TARGET .002 20:00
BLANK TEST .000 20:01
INTERNAL STANDARD VERIFIED 20:01
ALCO TV 30.07 in .002 20:01
BLANK TEST .000 20:02
SUBJECT SAMPLE *CBS 7/30/10* .000 20:03
BLANK TEST .000 20:03
ALCO TV 30.07 in .002 20:04
BLANK TEST .000 20:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

JULY 25, 2010
TIME 20:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 30.07 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! "# \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \ ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } + ~