

JAN 03 2010 ^{DNW} 3/4/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130584 ✓

Supervisor/Operator Performing the Verification Procedure:

Name Supt. James R. R. Rea

ID# 3161 ✓

Date 12-21-10 ✓

A Agency Kotzebue Police Department

Phone # 907-442-3352

Instrument Location Kotzebue Regional Jail 258d rm#107 Fourth Avenue Kotzebue, AK 99752

Alco S/N 77180 ✓

Target Value .079 ✓

High Pressure 1200>

B

Alco Test Values:

.080 ✓
1st Alco

.080 ✓
2nd Alco

Signature _____

Supt. James R. R. Rea
(OVER)

COB
1/18/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of MAR, 2010 ^{DNW}

(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130584

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

DECEMBER 21, 2010

OPERATOR'S NAME:

REAR/JAMES/RR

OPERATOR'S NUMBER: 3161

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

VERIFICATION OF CAL

O.L. #: 12212010

DEPT/AGENCY: OTZI

CASE/REPORT: 12212010

TEST TYPE: V

ALCO TARGET VALUE: 079

ALCO S/N: 77120

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.55 in
ALCO TARGET .000 19:06
BLANK TEST .000 19:07
INTERNAL STANDARD VERIFIED 19:07
ALCO TV 30.55 in .000 19:07
BLANK TEST .000 19:08
SUBJECT SAMPLE .000 19:09
BLANK TEST .000 19:10
ALCO TV 30.55 in .000 19:10
BLANK TEST .000 19:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130584

DECEMBER 21, 2010

TIME 19:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.55 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~