

APR 27 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130660</u>
Name	<u>DARYL RICE</u>	ID# <u>2088</u> Date <u>04/19/09</u>
A Agency	<u>SETKA POLICE DEPT.</u>	Phone # <u>747-3245</u>
Instrument Location <u>304 LAKE STREET</u>		
B Alco S/N <u>X124617</u> ✓ Target Value <u>.085</u> High Pressure <u>900</u>		
Alco Test Values		
	<u>.091</u> ✓ 1 st Alco	<u>.091</u> ✓ 2 nd Alco
Signature	<u>Daryl Rice</u>	<u>003</u> <u>5/13/09</u>

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May, 2009.

Carolyn M. Noland (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130660 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Daryl Rice
04/19/09
WCS PST
900

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130660

APRIL 19, 2009

OPERATOR'S NAME: *CS*
RICE DARYL DEAN
OPERATOR'S NUMBER: 2008
SUBJECT'S LAST NAME: *04/19/09*
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: SIT1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .085 ✓
ALCO S/N: X124617 ✓

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.35 in
ALCO TARGET .086 22:06
BLANK TEST .000 22:07
INTERNAL STANDARD VERIFIED 22:07
ALCO TV 30.35 in .091 ✓ 22:07
BLANK TEST .000 22:08
SUBJECT SAMPLE *CS* .000 22:08
BLANK TEST *04/19/09* .000 22:09
ALCO TV 30.33 in .091 ✓ 22:09
BLANK TEST .000 22:10

DIAGNOSTIC CHECK

Daryl Rice
04/19/09
WCS PST
900

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130660

APRIL 19, 2009 ✓
TIME 22:03

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 41c
BAROMETER: 30.37 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMN0PQRSTUWXYZ\]^_`ab cde fgh i j k l m n o
p q r s t u v w x y z { | } ~