

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

AUG 16 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130660</u> ✓	
Name <u>Garrett R. Willis</u>		ID# <u>5560</u> ✓	Date <u>8/13/10</u> ✓
A	Agency <u>AST</u>	Phone # <u>907-262-4453</u>	
Instrument Location <u>Cooper Landing</u>			
Alco S/N <u>57345</u> ✓		Target Value <u>.082</u> ✓	High Pressure <u>700</u>
B	Alco Test Value Average	<u>.086</u> ✓ 1 <sup>st</sup> Alco	<u>.086</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>Garrett Willis</u>			
(OVER)			

*BMB*  
*8/27/10*


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of Nov, 2010.

*Daniel L. Nicholas* (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130660

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130660

AUGUST 13, 2010

OPERATOR'S NAME:  
WILLIS/GARRETT/R  
OPERATOR'S NUMBER: 5560  
SUBJECT'S LAST NAME:  
VOC

SUBJECT'S FIRST NAME/MI :  
VOC

D.L. #: VOC  
DEPT/AGENCY: C000  
CASE/REPORT: VOC  
TEST TYPE: V

ALCO TARGET VALUE: .082  
ALCO S/N: 57345

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.01 in  
ALCO TARGET .082 14:54  
BLANK TEST .000 14:55  
INTERVAL STANDARD VERIFIED 14:55  
ALCO TV 29.99 in .086 14:55  
BLANK TEST .000 14:56  
SUBJECT SAMPLE .000 14:56  
BLANK TEST .000 14:57  
ALCO TV 30.01 in .086 14:57  
BLANK TEST .000 14:58

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130660

AUGUST 13, 2010  
TIME 14:52

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~