

OCT 25 2010

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130600 ✓

Name Garrett R. Willis

ID# 5560 ✓

Date 10/22/10 ✓

A

Agency AST

Phone # 362-4453

Instrument Location Cooper Landing AST

B

Alco S/N 51345 ✓

Target Value .082 ✓

High Pressure 650

Alco Test Value Average .083 ✓  
1<sup>st</sup> Alco

.083 ✓  
2<sup>nd</sup> Alco

Signature Garrett Willis

(OVER)

*OMB*  
*12/7/10*


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 7<sup>th</sup> day of DEC, 2010.

*Daniel L. Nicholas*  
(Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130660

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130660

OCTOBER 22, 2010

OPERATOR'S NAME:

WILLIS/GARRETT/R

OPERATOR'S NUMBER: 5560

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: V

DEPT/AGENCY: C000

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .082

ALCO S/N: 57345

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.26 in  
ALCO TARGET .000 14:34  
BLANK TEST .000 14:35  
INTERNAL STANDARD VERIFIED 14:35  
ALCO TV 29.26 in .083 14:35  
BLANK TEST .000 14:36  
SUBJECT SAMPLE .000 14:36  
BLANK TEST .000 14:37  
ALCO TV 29.26 in .083 14:37  
BLANK TEST .000 14:38

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130660

OCTOBER 22, 2010

TIME 14:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.28 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrs tuvwxyz{|}~