

VERIFICATION OF CALIBRATION REPORT


of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 29 2011


APR 08 2011

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666 ✓			
Name	John Merrion	ID#	4785 ✓			
		Date	3/18/11 ✓			
A	Agency	Fairbanks Police Department	Phone # 450-6500			
	Instrument Location	911 Cushman Street Fairbanks, Alaska 99701				
B	Alco S/N	X124626 ✓	Target Value	.077 ✓	High Pressure	1000
	Alco Test Value Average	.075 ✓		.075 ✓		
		1 st Alco		2 nd Alco		
	Signature					
		(OVER)				

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26 day of MAY, 2011.

 (Notary Seal Stamp)
Daniel L. Nicholas KYLE KLOPPER
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

MARCH 18, 2011

OPERATOR'S NAME: MERRION, JOHN/E
OPERATOR'S NUMBER: 4785
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: FBK1
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: X124626

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.86 in
ALCO TARGET .076 16:41
BLANK TEST .000 16:42
INTERNAL STANDARD VERIFIED 16:42
ALCO TV 29.86 in .075 16:43
BLANK TEST .000 16:43
SUBJECT SAMPLE .000 16:44
BLANK TEST .000 16:45
ALCO TV 29.86 in .075 16:45
BLANK TEST .000 16:46

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

MARCH 18, 2011
TIME 16:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.86 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./0123456789:;<=?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~