

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 04 2008

|   |                      |                                  |                          |
|---|----------------------|----------------------------------|--------------------------|
| Supervisor/Operator Performing the Verification Procedure:    |                      | DataMaster cdm S/N <u>130660</u> |                          |
| Name <u>JOHN MERRION</u>                                      |                      | ID# <u>4785</u>                  | Date <u>01/26/08</u>     |
| A   | Agency <u>FPD</u>    | Phone # <u>450-6500</u>          |                          |
| Instrument Location <u>FAIRBANKS POLICE DEPT. 911 CUSHMAN</u> |                      |                                  |                          |
| Alco S/N <u>X301542</u>                                       |                      | Target Value <u>.084</u>         | High Pressure <u>950</u> |
| B   | Alco Test Values     |                                  |                          |
|   | <u>.085</u>          | <u>.084</u>                      |                          |
|   | 1 <sup>st</sup> Alco | 2 <sup>nd</sup> Alco             |                          |
| Signature <u>[Signature]</u>                                  |                      | <u>CSB</u><br><u>2/4/08</u>      |                          |
| (OVER)  |                      |                                  |                          |

Under the Alaska Rules of Evidence, I certify that <sup>(If not, write in the area below)</sup>

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Merrion, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

2/8/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

JANUARY 26, 2008 *CSB 2/2/08*

OPERATOR'S NAME:  
HERRION/JOHN/E  
OPERATOR'S NUMBER: 4785  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :

A  
O.L. #: A  
DEPT/AGENCY: FBK1  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: .084  
ALCO S/N: X301542

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.37 in  
ALCO TARGET .085 19:47  
BLANK TEST .000 19:47  
INTERNAL STANDARD VERIFIED 19:48  
ALCO TV 30.37 in .085 19:48  
BLANK TEST .000 19:49  
SUBJECT SAMPLE .000 *CSB 2/2/08* 19:49  
BLANK TEST .000 19:50  
ALCO TV 30.37 in .084 19:50  
BLANK TEST .000 19:51

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

JANUARY 26, 2008  
TIME 19:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.37 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z ( ) \* +