

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUN 22 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666
Name		Gary Yamamoto	ID# 2765 Date 6-18-09
A	Agency	Fairbanks Police Department	Phone # 459-6500
Instrument Location 911 Cushman Street Fairbanks, Alaska 99701			
Alco S/N		58442	Target Value .079 High Pressure 1050
B	Alco Test Value Average	.082	.082
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<i>[Handwritten Signature]</i>	<i>[Handwritten Initials]</i> 6/25/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

*[Handwritten Signature]*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of June 2009.

*[Handwritten Signature]*  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

JUNE 18, 2009

OPERATOR'S NAME:

YAMAMOTO

OPERATOR'S NUMBER: 2765

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: 1111

DEPT/AGENCY: FBK1

CASE/REPORT: 1111

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 59442 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.49 in ✓

ALCO TARGET	.077	19:12
BLANK TEST	.000	19:13
INTERNAL STANDARD	VERIFIED	19:13
ALCO TV 29.49 in	.082 ✓	19:13
BLANK TEST	.000	19:14
SUBJECT SAMPLE	.000	19:15
BLANK TEST	.000	19:15
ALCO TV 29.49 in	.082 ✓	19:16
BLANK TEST	.000	19:16

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

JUNE 18, 2009 ✓

TIME 19:08

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.49 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""##%'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~