VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

Supervisor/Operator Performing the Verification Procedure:

Name Matthew L. Soden ID# 3378 Date 09-16-2008

A

Agency Fairbanks Police Department Phone # (907) 459-6500

Instrument Location 911 Cushman St. Fairbanks, AK 99701

B

Alco S/N X124611 Target Value .081 High Pressure 300 psi

Alco Test Values .082 (1st Alco) .083 (2nd Alco)

Signature CDB 9/22/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed.

(c) Instrument operator, Matthew Soden, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

10/15/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9-03/00

SEP 22 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 16, 2008 CDB 9/22/08

OPERATOR'S NAME:
SODEN, MATTHEW L
OPERATOR'S NUMBER: 3378
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
X
O.L. #: X
DEPT/AGENCY: FBK1
CASE/REPORT: X
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X124611

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.60 in

ALCO TARGET	.080	13:05
BLANK TEST	.000	13:06
INTERNAL STANDARD	VERIFIED	13:06
ALCO TV 29.60 in	.082	13:06
BLANK TEST	.000	13:07
SUBJECT SAMPLE	.000	13:08
BLANK TEST	.000	13:09
ALCO TV 29.60 in	.083	13:09
BLANK TEST	.000	13:10

CDB 9/22/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 16, 2008
TIME 12:49

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~