

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

SEP 25 2008

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666 ✓
Name		Gary Yamamoto	ID# 2765 Date 9-18-08 ✓
A	Agency	Fairbanks Police Department	Phone # 459-6500
Instrument Location		911 Cushman Street Fairbanks, Alaska 99701 ✓	
Alco S/N		X124611 ✓	Target Value .081 ✓ High Pressure 0
B	Alco Test Value Average	.083 ✓ 1 st Alco	.082 ✓ 2 nd Alco
Signature		 ✓	003 9/20/08


(OVER)

Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary Yamamoto, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 18, 2008

OPERATOR'S NAME:
YAMAMOTO
OPERATOR'S NUMBER: 2765
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 1111
DEPT/AGENCY: FBK1
CASE/REPORT: 1111
TEST TYPE: V
ALCO TARGET VALUE: .001 ✓
ALCO S/N: X124611

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.51 in
ALCO TARGET .079 01:43
BLANK TEST .000 01:43
INTERNAL STANDARD VERIFIED 01:44
ALCO TV 29.51 in .003 ✓ 01:44
BLANK TEST .000 01:45
SUBJECT SAMPLE :.000 01:45
BLANK TEST .000 01:46
ALCO TV 29.51 in .002 ✓ 01:46
BLANK TEST .000 01:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 18, 2008 ✓
TIME 01:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~