


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

SEP 25 2008

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666
Name		Gary Yamamoto	ID# 2765 Date 09-21-08 ✓
A	Agency	Fairbanks Police Department	Phone # 459-6500
Instrument Location		911 Cushman Street Fairbanks, Alaska 99701	✓
Alco S/N		X301737 ✓	Target Value .080 ✓ High Pressure 300
B	Alco Test Value Average	.083 ✓	.083 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature			
(OVER)			


Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

(a) I am Jeane Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Gary Yamamoto, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeane Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

10/24/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

SEPTEMBER 21, 2008 *CS 10/21/08*

OPERATOR'S NAME:

YAMAMOTO *Y*

OPERATOR'S NUMBER: 2785

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1111

DEPT/AGENCY: FBK1

CASE/REPORT: 1111

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X301737 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.75 in  
ALCO TARGET .079 05:12  
BLANK TEST .000 05:13  
INTERNAL STANDARD VERIFIED 05:13  
ALCO TV 29.75 in .003 ✓ 05:13  
BLANK TEST .000 05:14  
SUBJECT SAMPLE .000 *CS 10/21/08* 05:14  
BLANK TEST .000 05:15  
ALCO TV 29.75 in .003 05:15  
BLANK TEST .000 05:16

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

SEPTEMBER 21, 2008 ✓ *[Signature]*

TIME 05:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.75 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~