


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130666</u> ✓
Name	<u>Matthew L. Soden</u>	ID#	<u>3378</u>
		Date	<u>09-25-2008</u> ✓
A Agency		<u>Fairbanks Police Department</u>	Phone # <u>(907) 459-6500</u>
Instrument Location		<u>911 Cushman St. Fairbanks, AK 99701</u> ✓	
B Alco S/N		<u>X301563</u> ✓	Target Value <u>.077</u> ✓
		High Pressure	<u>900 psi</u>
Alco Test Values		<u>.082</u> ✓	<u>.081</u> ✓
		1 st Alco	2 nd Alco
Signature		 ✓	
		COS 10/24/08	
(OVER)			

Under the Alaska Rules of Evidence, I certify that (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Matthew Soden, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

10/24/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 25, 2000

OPERATOR'S NAME:

SODEN, MATTHEW L

OPERATOR'S NUMBER: 3378

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: FBK1

CASE/REPORT: X

TEST TYPE: V

ALCO TARGET VALUE: .077 ✓

ALCO S/N: X301563

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.78 in
ALCO TARGET .076 17:08
BLANK TEST .000 17:09
INTERNAL STANDARD VERIFIED 17:09
ALCO TV 29.78 in .002 ✓ 17:09
BLANK TEST .000 17:10
SUBJECT SAMPLE .000 ✓ 17:11
BLANK TEST .000 17:11
ALCO TV 29.78 in .081 ✓ 17:12
BLANK TEST .000 17:13

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

SEPTEMBER 25, 2000 ✓

TIME 17:01

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.78 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy{|}~