



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

OCT 22 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130666
Name	Gary Yamamoto	ID#	2765
Date	10-14-2009		
A	Agency	Fairbanks Police Department	Phone # 459-6500
Instrument Location	911 Cushman Street Fairbanks, Alaska 99701		
B	Alco S/N	X 301605	Target Value .086
			High Pressure 1300
Alco Test Value Average	.086	.086	
	1 st Alco	2 nd Alco	
Signature			
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of NOV, 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

