VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

A

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130666

Name JOHN MERRION ID# 4785 Date 12/05/08

Agency FAIRBANKS PD Phone # 450-6500

Instrument Location FPD

B

Alco S/N 78966 Target Value .077 High Pressure

Alco Test Values .078 (1st Alco) .078 (2nd Alco)

Signature

CDB 12/23/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Merrion , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

12/23/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

DEC 06/03 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

DECEMBER 05, 2008 CJB 12/23/08

OPERATOR'S NAME:
MERRION/JOHN/E
OPERATOR'S NUMBER: 4785
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L
O.L. #: L
DEPT/AGENCY: FBK1
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .077
ALCO S/N: 78966

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.60 in

ALCO TARGET	.076	21:25
BLANK TEST	.000	21:26
INTERNAL STANDARD	VERIFIED	21:26
ALCO TV 29.60 in	.078	21:27
BLANK TEST	.000	21:28
SUBJECT SAMPLE CJB 12/23/08	.000	21:28
BLANK TEST	.000	21:29
ALCO TV 29.60 in	.078	21:29
BLANK TEST	.000	21:30

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130666

DECEMBER 05, 2008
TIME 21:15

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	4c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~