

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130188</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>HUTT, William T.</u>	ID# <u>3359</u> Date <u>01/08/09</u>
A Agency <u>AST/Anchor Point Post (ANPE)</u>	Phone # <u>235-8239</u> <u>235-3000</u>
Instrument Location <u>34115 Sterling Hwy./P.O. Box 807</u> <u>Anchor Point,</u> <u>AK 99556</u> ✓	
B Alco S/N <u>X 301536</u> ✓ Target Value <u>.075</u> ✓ High Pressure <u>400 psi</u>	
Alco Test Values <u>.076</u> ✓ <u>.077</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>William T. Hutt</u> ✓ CSB 1/20/09	
(OVER)	

Under the Alaska Rules of Evidence, I certify that;
 (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Hutt, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz 2/3/09
 Jeanne Swartz Date
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

JAN 20 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188

JANUARY 08, 2009 *COB*

OPERATOR'S NAME:

HUTT, WILLIAM T

OPERATOR'S NUMBER: 3359

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

ETOH

O.I.L. #: 01234567

DEPT/AGENCY: AST0 / ANPE

CASE/REPORT: 09-00000

TEST TYPE: U

ALCO TARGET VALUE: .075 ✓

ALCO S/N: X301536

— BREATH ANALYSIS —

.075 ADJUSTED FOR 29.75 in

ALCO TARGET	.074	18:07
BLANK TEST	.000	18:08
INTERNAL STANDARD	VERIFIED	18:08
ALCO TU 29.75 in	.076 ✓	18:09
BLANK TEST	.000	18:09
SUBJECT SAMPLE <i>WTHD</i>	.000	18:10
BLANK TEST <i>WTHD</i>	.000	18:11
ALCO TU 29.75 in	.077 ✓	18:11
BLANK TEST	.000	18:12

DIAGNOSTIC CHECK.....

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188

JANUARY 08, 2009 ✓ *WTHD*
TIME 17:59

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.75 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢