

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 10 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130196</u> ✓
Name	<u>J Swartz</u>	ID# <u>4314</u> Date <u>5/11/09</u> ✓
A Agency	<u>SCOL</u>	Phone # <u>269-5592</u>
Instrument Location	<u>SCOL</u> ✓	
Alco S/N <u>SCOTTOR</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>750 psi</u>		
B Alco Test Values	<u>.082</u> ✓ 1 <sup>st</sup> Alco	<u>.082</u> ✓ 2 <sup>nd</sup> Alco
Signature	<u>[Signature]</u> ✓	
(OVER)		<u>BMB</u> <u>9/10/09</u>

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10<sup>th</sup> day of Oct, 2009.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130196

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

MAY 11, 2009

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: AAA1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: SCOTT08

### — BREATH ANALYSIS —

.000 ADJUSTED FOR 30.09 in ✓  
ALCO TARGET .000 14:37  
BLANK TEST .000 14:38  
INTERNAL STANDARD VERIFIED 14:38  
ALCO TV 30.09 in .002 14:38  
BLANK TEST .000 14:39  
SUBJECT SAMPLE .000 14:39  
BLANK TEST .000 14:40  
ALCO TV 30.09 in .002 14:40  
BLANK TEST .000 14:41

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130196

MAY 11, 2009

TIME 14:35

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 42c  
BAROMETER: 30.09 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ\* abcdefghijklmnop  
qrstuvwxyz{|}~