

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130287</u> ✓	
Name <u>Mitchell Langseth</u>		ID# <u>2612</u>	Date <u>12/30/08</u> ✓
Agency <u>Kenai Police Dept</u>		Phone # <u>283-7879</u>	
Instrument Location <u>Wildwood Prefab</u> ✓			
Alco S/N <u>X124627</u> ✓		Target Value <u>.079</u> ✓	High Pressure <u>500</u>
Alco Test Value Average <u>.081</u> ✓		<u>.081</u> ✓	
1 st Alco		2 nd Alco	
Signature <u>[Signature]</u> ✓		<u>008</u> <u>1/6/09</u>	

(OVER)


Under the Alaska Rules of Evidence, I certify that (write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Mitch E Langseth, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

1/6/09
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

DECEMBER 30, 2008

OPERATOR'S NAME: *cab 11/10/08*
LANGSETH

OPERATOR'S NUMBER: 2612

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

VERIFICATION

D.L. #: 0

DEPT/AGENCY: KENI

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124627

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.11 in		
ALCO TARGET	.079	09:37
BLANK TEST	.000	09:38
INTERNAL STANDARD	VERIFIED	09:38
ALCO TV 30.11 in	.081 ✓	09:39
BLANK TEST	.000	09:40
SUBJECT SAMPLE	.000	09:40
BLANK TEST	.000	09:41
ALCO TV 30.11 in	.081 ✓	09:42
BLANK TEST	.000	09:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130287

DECEMBER 30, 2008 ✓

TIME 09:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 43c

BAROMETER: 30.11 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~#