

SEP 13 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130113
Name		Kenneth G. Cox	ID# 3527 Date 9/5/2011
A	Agency	Skagway Police Department	Phone # (907) 983-2232
Instrument Location		Skagway Police Department 1st and State Street Skagway, AK 99840	
Alco S/N		X173019	Target Value .082 High Pressure 500 lbs
Alco Test Values		.078 1 st Alco	.079 2 nd Alco
Signature			


(OVER)

BMB
9/13/11


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 7 day of Nov, 2011.


 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

SEPTEMBER 05, 2011

OPERATOR'S NAME: *BMB 9/13/11*
COM/KENNETH/C
OPERATOR'S PHONE: 000
SUBJECT'S NAME: 000
IDENTIFICATION #/CAL: 000
SUBJECT'S FIRST NAME/MI: L
D.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2011—
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: X173019

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.72 in
ALCO TARGET .081 02:01
BLANK TEST .000 02:02
INTERNAL STANDARD .000 02:02
ALCO TU 29.73 in .082 02:03
BLANK TEST .000 02:03
SUBJECT SAMPLE .000 02:05
BLANK TEST .000 02:06
ALCO TU 29.73 in .079 02:06
BLANK TEST .000 02:07

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

SEPTEMBER 05, 2011
TIME 01:51

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/09/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 41c
BAROMETER: 29.72 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

0123456789:;<=>@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abdefghijklmno
pqrstuvwxyz{|}~