

APR 06 2011

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130113
A	Name	Kenneth G. Cox	ID# 3527 Date 4/3/2011
	Agency	Skagway Police Department	Phone # (907) 983-2232
	Instrument Location	Skagway Police Department 1st and State Street Skagway, AK 99840	
B	Alco S/N	X173019	Target Value .082 High Pressure 900
	Alco Test Values	.078 1st Alco	.079 2 nd Alco
	Signature	<i>Kenneth G. Cox</i>	

(OVER)

*BMB
4/7/11*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010 2011

NJB

Daniel L. Nicholas
 Daniel L. Nicholas *Kyle Klopfer*
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 03, 2011

OPERATOR'S NAME: COX, KENNETH G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME: L
O.L. # 1
DEPT/AGENCY: 2861
CASE REPORT: 2011-0523
TEST TYPE: U
ALCO TARGET VALUE: .002
ALCO S/N: K173019

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.54 in
ALCO TARGET .000 22:48
BLANK TEST .000 22:49
INTERNAL STANDARD VERIFIED 22:49
ALCO TO 29.54 in .078 22:50
BLANK TEST .000 22:51
SUBJECT SAMPLE .000 22:51
BLANK TEST .000 22:52
ALCO TO 29.54 in .079 22:53
BLANK TEST .000 22:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

APRIL 03, 2011
TIME 22:43

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.54 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*~.-/0123456789:;<=>?@ABCDEFG
HIJKLMNQPQRSTUVWXYZ[\]^_`ab cdefghijklmno
pqrstuvwxy{|}~#

