


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2011


Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130113</u>	
Name	<u>Kenneth G. Cox</u> ID# <u>3527</u> Date <u>July 6, 2011</u>
A Agency	<u>Skagway Police Department</u> Phone # <u>(907) 983-2232</u>
Instrument Location <u>Skagway Police Department 1st and State Street Skagway, AK 99840</u>	
Alco S/N <u>X173019</u> Target Value <u>.082</u> High Pressure <u>600 lbs</u>	
B Alco Test Values	<u>.080</u> <u>.081</u>
	1st Alco 2 nd Alco
Signature	
(OVER)	

Handwritten initials: BMB
7/27/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2010.

Handwritten initials: NJB
2011


(Notary Seal Stamp)
Daniel L. Nicholas NIKKI ROTH
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 06, 2011

OPERATOR'S NAME:
COV/KENNETH/G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L

O.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2011-2467
TEST TYPE: U
ALCO TARGET VALUE: .002
ALCO S/N: X173019

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 30.25 in
ALCO TARGET .002 00:28
BLANK TEST .000 00:29
INTERNAL STANDARD VERIFIED 00:29
ALCO TV 30.25 in .000 00:30
BLANK TEST .000 00:31
SUBJECT SAMPLE .000 00:32
BLANK TEST .000 00:33
ALCO TV 30.25 in .001 00:34
BLANK TEST .000 00:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

JULY 06, 2011
TIME 00:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.25 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~

