

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130114</u>	
Name <u>J Swartz</u>		ID# <u>4314</u>	Date <u>6/27/08</u> ✓
A	Agency <u>J Swartz</u> ✓	Phone # _____	
Instrument Location <u>SC DL</u>			
Alco S/N <u>SC07107</u> ✓		Target Value <u>.080</u>	High Pressure <u>400 psi</u>
B	Alco Test Values	<u>.078</u> 1 st Alco	<u>.077</u> 2 nd Alco
Signature <u>J Swartz</u>		<u>CSB</u> <u>7/3/08</u>	
(OVER)			


Under the Alaska Rules of Evidence, I certify that (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



J Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/7/08

Date

VERIFICATION OF CALIBRATION REPORT


of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130114

JUNE 27, 2008 ✓

Breath Test Record			
Serial Number	130114	Model	DataMaster
Location	ANCI	Department	Anchorage Police Dept.
Subject Information			
Last Name	Verification Of Cal	O.L. #	L
First Name	L		
Operator Information			
Operator Name	SwartzJ	Case	
Operator #	4314	Department	AAA1
Test Results			
Alco Target	.080	Alco Serial	SCOTT07
Nominal	0.079	Barometer	29.70
Dry Gas 1	0.078	Time	11:31:19
Subject Alcohol	0.000	Time	11:32:44
Dry Gas 2	0.077	Time	11:33:53
		Test Date	06/27/2008
		Test Time	11:29:56
		Report	<input checked="" type="checkbox"/>

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130114

JUNE 27, 2008 ✓
 TIME 11:28

--- DIAGNOSTIC CHECK ---

Diagnostic Test Record	
Serial Number	130114
Model	DataMaster
Location	ANCI
Department	Anchorage Police Dept.
Diagnostic Test Results	
Test Date	06/27/2008
Test Time	11:28:23
Computer	Okay
Program	Okay
Software Date	02/20/2001
Sample Chamber (C)	48
Breath Tube (C)	42
Barometer (in)	29.70
Flow Detector	Okay
High Speed Pump	Okay
Detector	Okay
Filter	Okay
Quartz	Okay
Calibration	Okay

Under the authority of the Alaska Department of Public Safety, I, the undersigned Director, do hereby certify that:

(a) I am the Scientific Director of the State of Alaska Scientific Crime Detection Laboratory, whose instrument was made available to the State of Alaska Scientific Crime Detection Laboratory for the purpose of testing the instrument.

(b) The instrument was tested and found to be in compliance with the requirements of the State of Alaska Scientific Crime Detection Laboratory.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

ALCO TV 29.70 in	.078	11:31
BLANK TEST	.000	11:32
SUBJECT SAMPLE	.000	11:32
BLANK TEST	.000	11:33
ALCO TV 29.70 in	.077	11:33
BLANK TEST	.000	11:34

Signed and sealed,
 J 6/27/08

FILTERS: UNIT
 QUARTZ STANDARD
 CALIBRATION: OKAY
 PRINTER TEST
 !*#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
 MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
 pqrstuvwxyz{|}~#

Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

Date