

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130114 ✓

Name Kevin Stevenson ID# 2414 Date 07-23-09 ✓

A Agency WAA Police Dept Phone # 907-786-1120

Instrument Location WAA Police Dept ✓

B Alco S/N 75470 ✓ Target Value .080 ✓ High Pressure 900

Alco Test Value Average .081 ✓

1st Alco .081 ✓
2nd Alco

Signature [Handwritten Signature]

(OVER)

CSB
7/28/09


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of Aug 2009.

[Handwritten Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

JULY 23, 2009

OPERATOR'S NAME: STEVENSON KEVIN L
OPERATOR'S NUMBER: 2414
SUBJECT'S LAST NAME: VERIFICATION
SUBJECT'S FIRST NAME/MI: VERIFICATION

O.L. #: 21234
DEPT/AGENCY: ANK7
CASE/REPORT: 097777
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: 75470 ✓

— BREATH ANALYSIS —

.000 ADJUSTED FOR 29.77 in		
ALCO TARGET	.079	21:03
BLANK TEST	.000	21:04
INTERNAL STANDARD	VERIFIED	21:04
ALCO TU 29.77 in	.081	21:05
BLANK TEST	.000	21:06
SUBJECT SAMPLE	.000	21:06
BLANK TEST	.000	21:07
ALCO TU 29.77 in	.081	21:07
BLANK TEST	.000	21:08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

JULY 23, 2009
TIME 21:01

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 42c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~@