

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 01 2011

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130116

Name STEVE KEVAN ID# 6585 Date 8/23/11

A Agency ALASKA STATE TROOPERS Phone # 907-438-6500

Instrument Location EMMONS AK POST

B Alco S/N 77111 Target Value 1078 High Pressure SSD

Alco Test Values 1074 1074
 1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
9/2/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2011.

[Signature] (Notary Seal Stamp)
 Carolyn M. Noland NIKKI ROTH
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130116

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

AUGUST 23, 2011

OPERATOR'S NAME:

KEVAN, STEVE

OPERATOR'S NUMBER: 6585

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 00

DEPT/AGENCY: EMMO

CASE/REPORT: 11-

TEST TYPE: V

ALCO TARGET VALUE: .078

ALCO S/N: 77111

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.57 in
ALCO TARGET .077 10:01
BLANK TEST .000 10:02
INTERNAL STANDARD VERIFIED 10:02
ALCO TV 29.57 in .074 10:02
BLANK TEST .000 10:03
SUBJECT SAMPLE .000 10:03
BLANK TEST .000 10:04
ALCO TV 29.57 in .074 10:05
BLANK TEST .000 10:06

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

AUGUST 23, 2011

TIME 09:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.57 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * +