



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

FEBRUARY 15, 2008

OPERATOR'S NAME: **COB 2/15/08**  
VALERIO, KYLE D  
OPERATOR'S NUMBER: 2093  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: XXXXXX  
DEPT/AGENCY: KOD1  
CASE/REPORT: 2008000000  
TEST TYPE: V  
ALCO TARGET VALUE: .002  
ALCO S/N: X124625

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.85 in  
ALCO TARGET .001 07:29  
BLANK TEST .000 07:30  
INTERNAL STANDARD VERIFIED 07:30  
ALCO TV 29.85 in .075 07:31  
BLANK TEST .000 **COB 2/15/08** 07:32  
SUBJECT SAMPLE .000 **COB 2/15/08** 07:32  
BLANK TEST .000 07:33  
ALCO TV 29.85 in .074 07:33  
BLANK TEST .000 07:35

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

FEBRUARY 15, 2008  
TIME 07:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 42c  
BREATH TUBE: 42c  
BAROMETER: 29.86 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJKL  
MNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~