

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

NOV 23 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130117</u> ✓	
Name <u>KYLE VALERIO</u>	ID# <u>2093</u> Date <u>11/19/2009</u> ✓
A Agency <u>KODIAK POLICE DEPARTMENT</u>	Phone # <u>(907)486-8000</u>
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK. 99615</u> ✓	
Alco S/N <u>X172938</u> ✓ Target Value <u>.087</u> ✓ High Pressure <u>900</u>	
B Alco Test Values	<u>.082</u> ✓ <u>.082</u> ✓ 1 <sup>st</sup> Alco                      2 <sup>nd</sup> Alco
Signature <u>[Handwritten Signature]</u>	
(OVER)	

BMB  
12/31/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 31<sup>st</sup> day of Dec, 2009.

[Handwritten Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



