

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 05 2010
JAN 05 2010
CIV

DataMaster cdm S/N 130117

Supervisor/Operator Performing the Verification Procedure:

Name KYLE VALERIO ID# 2093 Date 12/30/2009

A Agency KODIAK POLICE DEPARTMENT Phone # (907)486-8000

Instrument Location 217 LOWER MILL BAY ROAD, KODIAK, AK. 99615

B Alco S/N X172938 Target Value .087 High Pressure 800

Alco Test Values .083 .084
 1st Alco 2nd Alco

Signature [Signature] CDB
1/9/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of January, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117 ✓
DECEMBER 30, 2003 ✓ *CSB 1/11/0*

OPERATOR'S NAME:
WALERIO, KYLE D
OPERATOR'S NUMBER: 2053
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI :
TEST

O.L. #: XXXXXX
DEPT/AGENCY: K001
CASE/REPORT: 200990000
TEST TYPE: V
ALCO TARGET VALUE: .087 ✓
ALCO S/N: 0172275 ✓

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 30.04 in ✓

ALCO TARGET	.087	14:18
BLANK TEST	.000	14:19
INTERNAL STANDARD	VERIFIED	14:19
ALCO TV 30.04 in	.083 ✓	14:19
BLANK TEST	.000	14:20
SUBJECT SAMPLE	.000 ✓	14:21
BLANK TEST	.000	14:21
ALCO TV 30.04 in	.084 ✓	14:22
BLANK TEST	.000	14:22

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130117 ✓
DECEMBER 30, 2003 ✓
TIME 14:23

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/09/03
HEATERS
SAMPLE CHANGER: 40c
BREATH TUBE: 40c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH PRESS: 3.37
DETECTOR: OKAY
FILTERS: OKAY
DIAGNOSTIC STANDARD: OKAY
CALIBRATION: OKAY

PERFORM TEST
ALCO TV 30.04 in = .083 ✓
ALCO TV 30.04 in = .084 ✓
ALCO TV 30.04 in = .084 ✓