

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

SEP 22 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130119

Name Tobin P Brennan

ID# 4542 Date 9-19-09

A

Agency Soldotna Police Department

Phone # (907) 262-4455

Instrument Location Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)

B

Alco S/N X301573

Target Value .080

High Pressure 800

Alco Test Values

.079
1st Alco

.079
2nd Alco


Signature 

(OVER)

BMB
10/13/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 19th day of Oct, 2009.



Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

SEPTEMBER 19, 2009

OPERATOR'S NAME: BRENNAN
OPERATOR'S NUMBER: 4042
SUBJECT'S LAST NAME:
SUBJECT'S FIRST NAME:
SUBJECT'S BIRTH DATE:
SEX:
HT: 5' 10"
WT: 180
TEST NUMBER: 100
BLOOD TARGET VALUE: .088 ✓
ALCO S/N: K381573 ✓

--- BREATH ANALYSIS ---

.088 ADJUSTED FOR 29.54 in
BLOOD TARGET .078 00143
BLANK TEST .080 00144
INTERNAL STANDARD .080 00144
BLOOD TEST .080 00144 ✓
BLOOD TEST .080 00144 ✓

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

SEPTEMBER 19, 2009
S/N 130119

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03-03-08
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 43c
THERMISTOR: 29.54 in
FLAME DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

ALCO S/N: K381573
ALCO S/N: K381573
ALCO S/N: K381573