


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


NOV 20 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u> ✓
Name <u>Tobin P Brennan</u>		ID# <u>4542</u> Date <u>11-17-09</u> ✓ <i>Cdm</i>
A	Agency <u>Soldotna Police Department</u>	Phone # <u>(907) 262-4455</u>
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u> ✓		
Alco S/N <u>X301573</u> Target Value <u>.080</u> High Pressure <u>400</u>		
B	Alco Test Values	<u>.078</u> <u>.079</u> 1 st Alco 2 nd Alco
Signature 		
(OVER)		

BMB
12/31/09

- (Do Not write in the area below)
- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
 - (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Dec, 2009.

(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

NONDRINKING SUBJECT TEST

ALCOHOL TARGET VALUE: 0.000

BMB
12/31/09

BMB
12/31/09

DIAGNOSTIC CHECK

DIAGNOSTIC CHECK

ALCOHOL TARGET VALUE: 0.000

BMB
12/31/09