

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130119 ✓

Supervisor/Operator Performing the Verification Procedure:  
 Name Tabin P Brennan ID# 4542 Date 9-21-10 ✓

**A** Agency Soldotna Police Dept Phone # 262-4455  
 Instrument Location Soldotna Police

---

**B** Alco S/N X301536 ✓ Target Value .081 ✓ High Pressure 1600  
 Alco Test Values .083 ✓ .083 ✓  
                                     1st Alco                                      2nd Alco

Signature [Signature]

(OVER) BMB  
9/27/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

[Signature] (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130119

SEPTEMBER 21, 2010

OPERATOR'S NAME:

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

LOC

D.L. #: 100

DEPT/AGENCY: SLD1

BASE/REPORT: 140

TEST TYPE: U

ALCO TARGET VALUE: .081

ALCO S/N: K301576

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.83 in  
ALCO TARGET .000 09:07  
BLANK TEST .000 09:07  
INTERNAL STANDARD VERIFIED 09:07  
ALCO @ 29.83 in .083 09:08  
BLANK TEST .000 09:09  
SUBJECT SAMPLE .080 09:09  
BLANK TEST .000 09:10  
ALCO @ 29.83 in .083 09:10  
BLANK TEST .000 09:11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130119

SEPTEMBER 21, 2010

TIME 09:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 44c  
BAROMETER: 29.83 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

\*#kz'()\*+,-.0123456789:;<=/?@B^CDEF  
HIJKL\*OPQRSTUWVYZ[\]^\_`abode'ghijklmno  
rstuvwxyz{|}~