VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

Supervisor/Operator Performing the Verification Procedure:

A

Name Kyle Carrington ID# 5336 Date 12-6-08

Agency Fairbanks Airport Police Phone # 474-2555

Instrument Location 5195 Brumbaugh

B

Alco S/N 78895 Target Value .081 High Pressure 575 PSI

Alco Test Values .079 (1st Alco) .078 (2nd Alco)

Signature [signature] CAB 12/18/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 12/20/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/ DEC 11 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130121

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

DECEMBER 06, 2008

OPERATOR'S NAME: CB 12/18/08
CARRINGTON,KYLE
OPERATOR'S NUMBER: 5336 KC
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
I
O.L. #: 0
DEPT/AGENCY: FBK6
CASE/REPORT: 0 KC
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: 78895

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.39 in

ALCO TARGET	.079	10:20
BLANK TEST	.000	10:22
INTERNAL STANDARD	VERIFIED	10:22
ALCO TV 29.39 in	.079	10:22
BLANK TEST	.000	10:23
SUBJECT SAMPLE KC	.000	10:23
BLANK TEST CB	.000	10:24
ALCO TV 29.39 in 12/18/08	.078	10:25
BLANK TEST	.000	10:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130121

DECEMBER 06, 2008
TIME 10:17

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/29/00
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	44c
BAROMETER:	29.39 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~