

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 1 : 2011
 2010
 3/8/11

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130124</u> ✓	
Name <u>NICHOLAS YLIENTEMZ</u>	ID# <u>3429</u> ✓ Date <u>1/11/11</u> ✓
A Agency <u>METLAKATLA POLICE DEPT.</u> Phone # <u>907 886-4011</u>	
Instrument Location <u>METLAKATLA POLICE DEPT.</u>	
B Alco S/N <u>X172988</u> ✓ Target Value <u>.078</u> ✓ High Pressure <u>1200 psi</u>	
Alco Test Values <u>.079</u> ✓	<u>.079</u> ✓
1 st Alco	2 nd Alco
Signature <u>Nicholas Ylientemz</u>	
(OVER)	

amb
2/8/11


I, Nita J. Bolz, after being first duly sworn, depose and state as follows (in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 11th day of MAR, 2011

Daniel L. Nicholas (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130124

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124
JANUARY 11, 2011
OPERATOR'S NAME: VLINENTI/NICHOLAS/L
OPERATOR'S NUMBER: 3429
SUBJECT'S LAST NAME: UER
SUBJECT'S FIRST NAME/MI : OF CAL
O.L. #: 00000
DEPT/AGENCY: MET1
CASE/REPORT: 0000
TEST TYPE: U
ALCO TARGET VALUE: .078
ALCO S/N: X172908

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.33 in
ALCO TARGET .079 12:46
BLANK TEST .000 12:46
INTERNAL STANDARD VERIFIED 12:46
ALCO TV 30.32 in .079 12:47
BLANK TEST .000 12:48
SUBJECT SAMPLE .000 12:48
BLANK TEST .000 12:49
ALCO TV 30.33 in .079 12:49
BLANK TEST .000 12:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130124

JANUARY 11, 2011
TIME 12:44

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.32 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxy{|}~