

FEB 23 2011
2011
3/18/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130125 ✓

Name Anthony M. Wiles ID# 6877 ✓ Date 2/17/2011 ✓

A Agency AST Phone # 451-5100

Instrument Location 1979 Peger Rd. FBES, AK 99709

B Alco S/N 555124 ✓ Target Value .08090 ✓ High Pressure 600 psi

Alco Test Values .07890 ✓ .07890 ✓
1st Alco 2nd Alco

Signature Anthony Wiles

(OVER)

AMB
3/6/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of MAR, 2011.

Daniel L. Nicholas (Notary Seal Stamp)
Carolyn M. Notland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

FEBRUARY 17, 2011

OPERATOR'S NAME:

WILES/ANTHONY/MICHAEL

OPERATOR'S NUMBER: 6877

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

N

O.L. #: N

DEPT/AGENCY: FBK0

CASE/REPORT: N

TEST TYPE: U

ALCO TARGET VALUE: .000

ALCO S/N: 555124

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.99 in

ALCO TARGET	.000	10:59
BLANK TEST	.000	11:00
INTERNAL STANDARD	VERIFIED	11:00
ALCO TV 29.99 in	.078	11:00
BLANK TEST	.000	11:01
SUBJECT SAMPLE	.000	11:02
BLANK TEST	.000	11:02
ALCO TV 29.98 in	.078	11:03
BLANK TEST	.000	11:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

FEBRUARY 17, 2011

TIME 11:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.98 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () ~ |