VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

SEP 15 2009

Supervisor/Operator Performing the Verification Procedure: Datamaster cdm S/N 130125

A

Name GARY J. Tellep ID# 2281 Date 9-6-09

Agency AST Phone # 907-451 5100

Instrument Location AST HQ 1979 Peger Rd., FBKS. 99709

B

Alco S/N X301591 Target Value .083 High Pressure 800 psi

Alco Test Value Average .076 (1st Alco) .077 (2nd Alco)

Signature [signature] CAB 9/28/09

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Oct, 2009.

[signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 01/00

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory & Statewide Breath Alcohol Program

Datamaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C **TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES:**

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

SEPTEMBER 06, 2009 CDB 9/28/09

OPERATOR'S NAME:
TELLEP GARY J
OPERATOR'S NUMBER: 2281
SUBJECT'S LAST NAME:
VERIFICATIONOFCAL
SUBJECT'S FIRST NAME/MI :
N
O.L. #: N
DEPT/AGENCY: FBK0
CASE/REPORT: N
TEST TYPE: V
ALCO TARGET VALUE: .083
ALCO S/N: X301591

— BREATH ANALYSIS —

.083 ADJUSTED FOR 29.55 in

ALCO TARGET	.082	00:20
BLANK TEST	.000	00:21
INTERNAL STANDARD	VERIFIED	00:21
ALCO TV 29.55 in	.076	00:21
BLANK TEST	.000	00:22
SUBJECT SAMPLE	.000	00:23
BLANK TEST	.000	00:24
ALCO TV 29.55 in	.077	00:24
BLANK TEST	.000	00:25

CDB 9/28/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

SEPTEMBER 06, 2009
TIME 00:26

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.55 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~