

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130125</u> ✓	
Name <u>Michael A. Wery</u>		ID# <u>11-13-08 CHP</u> <u>3590</u>	Date <u>10-22-08</u> ✓
A	Agency <u>AST</u>	Phone # <u>451-5100</u>	
Instrument Location <u>Fairbanks</u> ✓			
Alco S/N <u>172942</u> ✓ Target Value <u>.077</u> ✓ High Pressure <u>850</u> ✓			
Alco Test Value Average			
1 <sup>st</sup> Alco <u>.075</u> ✓		2 <sup>nd</sup> Alco <u>.076</u> ✓	
Signature <u>Michael A. Wery</u> ✓			<u>CSB</u> <u>11/20/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Michael A Wery, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

11/21/08

Date

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130125 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125  
OCTOBER 22, 2008  
OPERATOR'S NAME: *CCB*  
VERA, MICHAEL A. *11/20/08*  
OPERATOR'S NUMBER: 3599  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/NI:  
OF CALIBRATION  
O.L. #: 0  
DEPT/AGENCY: FBK0  
CASE/REPORT:  
TEST TYPE: U  
ALCO TARGET VALUE: .077 ✓  
ALCO S/N: X172942

--- BREATH ANALYSIS ---

.077 ADJUSTED FOR 29.13 in		
ALCO TARGET	.074	11:44
BLANK TEST	.000	11:44
INTERNAL STANDARD	VERIFIED	11:44
ALCO TV 29.13 in	.075	11:45 ✓
BLANK TEST	.000	11:46
SUBJECT SAMPLE	.000	11:46
BLANK TEST	.000	11:47
ALCO TV 29.13 in	.076	11:47 ✓
BLANK TEST	.000	11:48

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130125  
OCTOBER 22, 2008 ✓  
TIME 11:42 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 29.13 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKL MNOPQRSTU VWXYZ \ ] ^ \_ ` abcde fgh i j k l m n o  
p q r s t u v w x y z { | } ~