

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | | | |
|--|--------------------------------------|------------------------------------|---------------------------|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N <u>130125</u> ✓ | |
| Name | <u>Michael A. Wery</u> | ID# | <u>3590</u> ✓ |
| | | Date | <u>6-14-10</u> ✓ |
| A | Agency <u>Alaska State Troopers</u> | Phone # | <u>451-2950</u> |
| | Instrument Location <u>Fairbanks</u> | | |
| B | Alco S/N <u>X 301536</u> ✓ | Target Value <u>.079</u> ✓ | High Pressure <u>1220</u> |
| | Alco Test Values | <u>.075</u> ✓ | <u>.075</u> ✓ |
| | | 1 st Alco | 2 nd Alco |
| | Signature | <u>[Signature]</u> | |

(OVER)

BMB
8/27/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Aug, 2010.

[Signature]
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



