

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130125</u>	
Name <u>Anthony M. WILES</u>	ID# <u>6877</u> Date <u>12/19/10</u>
Agency <u>AST</u>	Phone # <u>451-5100</u>
Instrument Location <u>1979 PEBER Rd. FBLS, AK 99709</u>	
Alco S/N <u>555124</u> Target Value <u>.08090</u> High Pressure <u>900 PSI</u>	
Alco Test Values	<u>.07790</u> <u>.07890</u>
1 st Alco	2 nd Alco
Signature <u>Anthony Wiles</u>	<u>COB</u> <u>12/30/10</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of FEB, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130125

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

DECEMBER 19, 2010

COB 12/20/10

OPERATOR'S NAME:
WILES/ANTHONY/MICHAEL
OPERATOR'S NUMBER: 6377
SUBJECT'S LAST NAME:
VERIFICATION/PCAL
SUBJECT'S FIRST NAME/MI :
N
O.L. #: N
DEPT/AGENCY: FBX0
CASE/REPORT: N
TEST TYPE: V
ALCO TARGET VALUE: .000
ALCO S/N: 555124

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.72 in
ALCO TARGET .079 11:59
BLANK TEST .000 11:00
INTERNAL STANDARD VERIFIED 11:00
ALCO TU 29.72 in .077 11:01
BLANK TEST .000 11:01
SUBJECT SAMPLE *COB 12/20/10* .000 11:02
BLANK TEST .000 11:03
ALCO TU 29.72 in .078 11:03
BLANK TEST .000 11:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

DECEMBER 19, 2010
TIME 11:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 41c
BAROMETER: 29.72 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!*"#&'()*+,-./0123456789:;<=>@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~