

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130126</u>	
Name: <u>Lt Eric Saffell</u>		ID#: <u>2853</u>	Date: <u>3/21/08</u> ✓
A	Agency: <u>Anchorage Airport Police</u> ✓	Phone #: <u>266-2579</u>	
Instrument Location: <u>Airport Police & Fire Station</u>			
Alco S/N <u>X301583</u> ✓ Target Value <u>.078</u> ✓ High Pressure <u>300 psi</u>			
B	Alco Test Values	<u>.076</u> ✓ 1 st Alco	<u>.075</u> ✓ 2 nd Alco
Signature: <u>Lt Eric Saffell</u>		<u>CSB</u> <u>3/24/08</u>	
(OVER)			

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Eric Saffell, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 3/24/08
Jeanne Swartz Date:
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130126

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 21, 2008

OPERATOR'S NAME: SFPELL, ERIC
OPERATOR'S NUMBER: 2853
SUSPECT'S LAST NAME:
VERIFICATION OF CAL
SUSPECT'S FIRST NAME/MI:

CAL. #: L
DEPT/AGENCY: ANCS
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: K301583

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.77 in
ALCO TARGET .077 02:39
BLANK TEST .000 02:40
INTERNAL STANDARD VERIFIED 02:40
ALCO TU 29.77 in .076 02:40
BLANK TEST .000 02:41
SUBJECT SAMPLE .000 02:41
BLANK TEST .000 02:42
ALCO TU 29.77 in .075 02:42
BLANK TEST .000 02:43

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

MARCH 21, 2008
TIME 02:35

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.77 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL^MNPQRSTUUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~