

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

OCT 20 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: 130126 ✓
A	Name: Lt. Eric Saffell ID#: 2853	Date: 10/20/09 ✓
	Agency: Anchorage Airport Police	Phone #: 266-2579
	Instrument Location: Anchorage Airport Police Department ✓	
B	Alco S/N: X172961 ✓	Target Value: .079 ✓
		High Pressure: 1100 psi
	Alco Test Value Average: .078 ✓	.077 ✓
	1 st Alco	2 nd Alco
	Signature: <u>Lt Eric Saffell</u>	
	(OVER)	

BMB
11/14/09

- I, Nita J. Bolz, after being first duly sworn, depose (and swear to) the following in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20 day of Dec, 2009.
Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

OCTOBER 28, 2009 ✓
TIME 20:27

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01
 HEATERS
 SAMPLE CHAMBER: 4%
 BREATH TUBE: 41c
 BAROMETER: 29.60 in
 FLOW DETECTOR: OKAY
 PUMP
 HIGH SPEED: OKAY
 DETECTOR: OKAY
 FILTERS: OKAY
 QUARTZ STANDARD: OKAY
 CALIBRATION: OKAY

PRINTER TEST

```
! " # $ % & ' ( ) * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z I \ ] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢
```

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

OCTOBER 28, 2009 ✓

OPERATOR'S NAME: SAFFELL/ERIC
 OPERATOR'S NUMBER: 2853
 SUBJECT'S LAST NAME:
 VERIFICATION OF CAL
 SUBJECT'S FIRST NAME/MI :
 L
 O.L. #: L
 DEPT/AGENCY: ANCS
 CASE/REPORT:
 TEST TYPE: U
 ALCO TARGET VALUE: .079 ✓
 ALCO S/N: X172961

BND
11/14/09

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.60 in		
ALCO TARGET	.080	20:32
BLANK TEST	.000	20:33
INTERNAL STANDARD	VERIFIED	20:33
ALCO TV 29.60 in	.078 ✓	20:33
BLANK TEST	.000	20:34
SUBJECT SAMPLE	.080	20:35
BLANK TEST	.000	20:36
ALCO TV 29.60 in	.077 ✓	20:36
BLANK TEST	.000	20:37

BND
11/14/09