

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130127

Name Christopher Umbs ID# 3669 Date 3/1/12

A Agency Alaska State Troopers Phone # 465-4004

Instrument Location Angoon

B Alco S/N 4Q Target Value .080 High Pressure 900psi

Alco Test Values .079 .079
1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
3/1/12

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012 (initials)
2010. 4/5/12

[Signature]
Nikki Roth, Notary Public
My Commission Expires With Office

BT9 06403



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130127

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127

MARCH 01, 2012

OPERATOR'S NAME: *BOMB 3/6/12*
LINDS/CHRISTOPHER/J
OPERATOR'S NUMBER: 3669
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/NI :
UOC

D.L. #: UOC
DEPT/AGENCY: ANSO
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .050
ALCO S/N: 40

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.94 in
ALCO TARGET .000 09:20
BLANK TEST .000 09:21
INTERNAL STANDARD VERIFIED 09:21
ALCO TV 29.94 in .079 09:21
BLANK TEST .000 09:22
SUBJECT SAMPLE *BOMB 3/6/12* .000 09:23
BLANK TEST .000 09:23
ALCO TV 29.94 in .079 09:24
BLANK TEST .000 09:24

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130127

MARCH 01, 2012
TIME 09:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~@