

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130128 ✓
Name	Byron D. Redburn	ID#	2087
		Date	4/23/2008 ✓
A	Agency	Nome Police Dept ✓	Phone # _____
	Instrument Location	Nome Police Dept	
B	Alco S/N	X124646 ✓	Target Value .080 ✓
		.0838 4/23/08	High Pressure 500
	Alco Test Values	.081 ✓	.083 ✓
		1 st Alco	2 nd Alco
Signature	J.P. Kella		CSB 4/28/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Byron D Redburn, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 4/29/08
Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130128

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130128

APRIL 23, 2008

OPERATOR'S NAME:
REDEGAN BYRON D
OPERATOR'S NUMBER: 2087
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

A

O.L. #: A
DEPT/AGENCY: OME1
CASE/REPORT: A
TEST TYPE: U
ALCO TARGET VALUE: .060
ALCO S/N: X124646

--- BREATH ANALYSIS ---

.060 ADJUSTED FOR 30.32 in
ALCO TARGET .061 09:17
BLANK TEST .000 09:17
INTERNAL STANDARD VERIFIED 09:18
ALCO TV 30.32 in .063 09:18
BLANK TEST .000 09:19
SUBJECT SAMPLE .000 09:19
BLANK TEST .000 09:20
ALCO TV 30.32 in .063 09:20
BLANK TEST .000 09:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130128

APRIL 23, 2008
TIME 09:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 30.32 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~