

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130129</u>
Name <u>George S. Kammer</u>		ID# <u>3840</u> Date <u>5-10-08</u> ✓
A	Agency <u>AST / UNALASKA</u> ✓	Phone # <u>624.3073</u>
Instrument Location <u>UNALASKA AST POST</u>		
Alco S/N <u>X301543</u> ✓		Target Value <u>.084</u> ✓ High Pressure <u>1000</u>
B	Alco Test Values	
	<u>.082</u> ✓ 1 st Alco	<u>.082</u> ✓ 2 nd Alco
Signature <u>[Signature]</u> <u>CSK</u>		<u>CSK</u> <u>6/18/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, George Kammer, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature] 6/19/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

*6567
5-10-08*

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

MAY 10, 2008 *6/18/08*

OPERATOR'S NAME:
KAMMER/GEORGE/J
OPERATOR'S NUMBER: 3840
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: UNKA
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .084 ✓
ALCO S/N: X301563 ✓

— BREATH ANALYSIS —

.084 ADJUSTED FOR 30.01 in		
ALCO TARGET	.084	00:19
BLANK TEST	.000	00:20
INTERNAL STANDARD	VERIFIED	00:21
ALCO TU 30.01 in	.082	00:21
BLANK TEST	.000	00:22
SUBJECT SAMPLE	.000 <i>6/18/08</i>	00:22
BLANK TEST	.000	00:23
ALCO TU 30.01 in	.082	00:23
BLANK TEST	.000	00:24

DIAGNOSTIC CHECK

*6567
5-17-08*

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

MAY 10, 2008 ✓
TIME 00:02

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.04 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~