

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130129

BMB
12/9/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

OCTOBER 28, 2010

OPERATOR'S NAME:

NICHOLS, JAMES/D

OPERATOR'S NUMBER: 6582

SUBJECT'S LAST NAME:

YOC

SUBJECT'S FIRST NAME/MI :

YOC

O.L. #: YOC

DEPT/AGENCY: BAWI

CASE/REPORT: YOC

TEST TYPE: V

ALCO TARGET VALUE: .004

ALCO S/N: X301553

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.57 in

ALCO TARGET .003 08:39

BLANK TEST .000 08:40

INTERNAL STANDARD VERIFIED 08:40

ALCO TV 29.57 in .002 08:40

BLANK TEST .000 08:41

SUBJECT SAMPLE .000 08:42

BLANK TEST .000 08:42

ALCO TV 29.57 in .001 08:43

BLANK TEST .000 08:43

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

OCTOBER 28, 2010

TIME 08:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.57 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD = OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ^_`abcde:ghijklmnop
qrs:tuvwxyz{|}~